
Development of a novel real-time method to study the erythrocytic life cycle of *Plasmodium falciparum* using Quartz Crystal Microbalances

Entwicklung einer neuen Echtzeit-Methode zur Untersuchung des erythrozytärer Lebenszyklus von *Plasmodium falciparum* mit Schwingquarzbiosensoren

DISSERTATION

der Fakultät für Chemie und Pharmazie
der Eberhard Karls Universität Tübingen

zur Erlangung des Grades eines Doktors
der Naturwissenschaften

2010

Vorgelegt von

ROSA ELENA BUSTOS CRUZ

Tag der mündlichen Prüfung:

25. Februar 2010

Dekan:

Professor Dr. L. Wesemann

1. Berichterstatter:

Professor Dr. H. Northoff

2. Berichterstatter:

Professor Dr. S. Laufer

This doctoral thesis was written during my work on 1.08.2006-15.02.2010 within the Biosensors Group (AG Gehring) of the Institute of Experimental and Clinical Transfusion Medicine under the guidance of Dr. rer. nat. Frank K. Gehring in cooperation with Dr. Mordmüller (Institut of Tropical Medicine, University of Tübingen). The evaluation was organized by Prof. Dr. Northoff, Institute of Clinical and Experimental Transfusion Medicine University of Tübingen, and Prof. Dr. Laufer of Institut of Pharmacy, University of Tübingen.

ABSTRACT

Development of a novel real-time method to study the erythrocytic life cycle of *Plasmodium falciparum* using Quartz Crystal Microbalances (QCM)

Previous studies carried out by the Biosensor Research Group of the Institute of Clinical and Experimental Transfusion Medicine at the Tübingen University had successfully demonstrated determination of various blood types through the interaction of erythrocytes with immobilised antibodies on a QCM. These studies have opened the possibility to test new applications of this technique for the study of infectious diseases. In our case, studies of the erythrocyte life cycle of *P. falciparum*, particularly during the last six hours preceding the merozoite release and studies related to reinfection of by merozoites were the focus of our investigation.

Release of malaria parasites of *P. falciparum* from infected erythrocyte at the end of their asexual erythrocyte cycle occurs approximately every 48 hours, asynchronously in parasite culture and synchronously *in vivo*. This process is poorly studied due to: (I) merozoite release is a very short event (<60 S), (II) schizont stages have high sensitivity for culture conditions like pH, medium, osmotic pressure, gas atmosphere and temperature. (III) Schizont conditions are not easy to maintain under the microscope, making microscopical real time observation of the release difficult.

The aim of the presented project was to study the release of *Plasmodium falciparum* merozoites from erythrocytes with the QCM sensor technique. In this way, the frequency shift due to the change of mass associated to the merozoite release and, on a second QCM, the reinfection of healthy erythrocytes are monitored in real time. Our QCM experiments included the following stages: (I) Adaptation and optimisation of the immobilisation of biological layers to capture the erythrocytes on the quartz. (II) Optimisation of parasite culture conditions in a QCM chamber. (III) Observation of the frequency signal both for infected and non-infected erythrocytes samples and correlation of the signal with the release of merozoites. (IV) Reinfection of healthy erythrocytes on a second quartz within one QCM system. (V) Test of inhibition of merozoite release and reinfection by antimalarial compounds.

The results showed that there was significant increase of ~1000 Hz for QCM with infected erythrocyte compared to QCM with healthy erythrocytes, where the frequency remained stable. Microscopical observation of the quartz surface at different times during the experiment (TEM and optical) demonstrated a correlation between the increase in frequency and merozoite release. At this point, approximately more than 80% of the infected erythrocytes on the quartz are involved in the release. Reinfection of new erythrocytes was observed on a second QCM. pH of the system (7.2), Temperature (37°C±0.1), flow of the medium (9 µL/min), sterility of the process (BactAlert), gas atmosphere (O₂ 5%, residual N₂) were established to ensure parasite development and survival. External controls using flow cytometry 24 hours after the reinfection show a parasitemia percentage of >1% in the erythrocytes infected *in situ*.

Our results show, that the QCM technique is an appropriate and important new tool to elucidate the biology of the re-invasion process of Plasmodia.

ACKNOWLEDGEMENT

I want to express my gratitude to all those who in so many different ways have helped me take this ship to shore. In particular, I would like to thank:

This thesis would not have been possible had it not been for the guidance, suggestions, support and encouragement provided by Prof. H. Northoff, chief of Institute of Clinical and Experimental Transfusion Medicine University of Tübingen. His understanding, encouragement and personal guidance have provided a good basis for the present thesis. I wish to express my gratitude to my second thesis advisor Prof. S. Laufer, director of the Pharmacy Institute University of Tübingen for the follow-up, care and suggestions kindly provided for my project.

I would like to thank Dr. F.K. Gehring, chief of the Biosensor Research Group for his kind guidance, valuable comments and encouragement. Their ideals and concepts have had a remarkable influence on my entire career in the field of Biosensors

I am grateful with Dr. B. Mordmüller of the Institute of Tropical Medicine, University of Tübingen, who introduced me to the field of Human parasitology. His encouragement, guidance and support from the initial to the final level enabled me to develop an understanding of the subject. I wish to extend my thanks to Serena, Jana, Rolph, Tanja of the human parasitology group for assisting me in many different ways

During this work I have collaborated with many colleagues for whom I have great regard, and I wish to extend my warmest thanks to all those who have helped me with my work in the Biosensor Research Group of Tübingen University

I am grateful to Prof. Schweizer and Mr. Jürgen of the Max Planck Institute for their support in Transmission Electron Microscopy.

This PhD work could not have been carried out without the support provided by the institution KAAD, who awarded me a scholarship for my doctoral stay in Germany. Similarly, their valuable assistance and support for me and my family were significant.

My sincere thanks are due to the official referees, Dr. E. Ninci M.D., and Dr. M. Desimonne, Ph.D, for their detailed review, constructive criticism and excellent advice during the preparation of this thesis.

I gratefully thank Dr. S. Quintero, PhD. for his constructive comments on this thesis.

I thank my friends in Tübingen, Hamburg and Colombia for their continuous moral support

I owe my deepest gratitude to my husband, Hugo Hernán, and my daughter, María Helena. Without their encouragement, tolerance, love and understanding, it would have been impossible for me to finish this work.

I cannot finish without saying how grateful I am with my family: grandparents, brothers, uncles, aunts, cousins and nephews who have given me a loving environment to develop.

Finally, I would like to thank everybody who was important to the successful realization of thesis, as well as to express my apology to those I could not mention personally one by one.

ABBREVIATIONS

ABRA	Acid Basic Repeat Antigen
AMA-	Apical Membrane Antigen
AO	Acridine Orange
ATCC	American Type Culture Collection
BAW	Bulk Acoustic Wave
CCD	Charge coupled device
CSA-	Chondroitin sulfate A
DARC	Duffy Antigen Receptor for Chemokines
DGs	Dense Granules
DHA	Dihydroartemisin
EBA-	Erythrocyte Binding Antigen
EPM	Erythrocyte Plasma Membrane
ER	Endoplasmic reticulum
FACS	Fluorescence Activated Cell Sorting
GA	Glutaraldehyde
GMEP	The Global Malaria Eradication Program
GPI-	Glycosylphosphatidylinositol
IC	Inhibitory concentration
ICAM-	Intracellular Adhesion Molecule
MACS	Magnetic Cell sorting
mH	Motional inductance
MR4	Malaria Research and Reference Reagent Resource Center
MSP-	Merozoite Surface Protein
MW	Molecular weight
PBS	Phosphate buffered saline
PF	Paraformaldehyde
Pf	Plasmodium falciparum
PfEMP-	<i>P. falciparum</i> Erythrocyte Membrane Protein
PfRBL	<i>P. falciparum</i> reticulocyte binding protein homolog (PfRh)
PLL	Poly-L-Lysine
PVM	Parasitophorus Vacuole Membrane
QCM	Quartz Crystal Microbalance
RAP	Rhoptry Associated Antigen
RBC	Red Blood Cells
RBCM	Red Blood Cell Membrane
RBL	Reticulocyte binding-like
RCD	Red Cell Deformability
RESA	Ring Erythrocyte Surface Antigen
Rhop	Rhoptries
RIMA	Ring Membrane Antigen
RPMI	Roswell Park Memorial Institute
RT	Room Temperature
SAMs	Self-assemble monolayers

SERA	Sera Repeat Antigen
SERCA	Sarco-endoplasmic reticulum Ca ⁺
SR	Sarcoplasmic reticulum
TBF	Thick Blood Film
TEM	Transmission Electron Microscopy
THBF	Thin Blood Film
TSMR	Thickness Shear Mode resonator
WHO	World Health Organisation

LIST OF FIGURES

- Figure 2-1. Principle of the function of a biosensor
- Figure 2-2. Mass-sensitive devices and acoustic wave propagation modes
- Figure 2-3. The piezoelectric effect
- Figure 2-4. The orientation of the polar axes of the crystal
- Figure 2-5. Cuts of α -quartz
- Figure 2-6. Geographical distribution of Malaria 2008
- Figure 2-7. Life cycle of *Plasmodium falciparum*
- Figure 2-8. Structure of Merozoite (*Plasmodium falciparum*)
- Figure 2-9. Merozoite invasion
- Figure 2-10. The asexual blood stages of *Plasmodium falciparum*
- Figure 2-11. Models for the release of *Plasmodium* merozoites from their host erythrocytes
- Figure 2-12. Receptor of *P. vivax* and *P. falciparum*
- Figure 2-13. Typical interfaces in biosensing
- Figure 2-14. Mechanism of oscillator biosensor
- Figure 3-1. Sensor platform Fidget-Type
- Figure 4-1. Schematic representation of the immobilisation of Poly-L-Lysine (PLL) over the gold electrode
- Figure 4-2. Frequency change of the resonance of Quartz coated with PLL
- Figure 4-3. Immobilised erythrocytes over the gold electrode
- Figure 4-4. Effect of incubation time and temperature
- Figure 4-5. Effect of washing and storage of the quartz before use
- Figure 4-6. Effect of regeneration
- Figure 4-7. Effect of molecular weight of PLL
- Figure 4-8. Comparison of the different films of PLL on gold electrode
- Figure 4-9. Graphic representation of the biosensor system used for the observation of the erythrocytes
- Figure 4-10. External observation of the erythrocytes on the quartz for 48 hours
- Figure 4-11. Microscopical optical pictures
- Figure 4-12. Observation of the frequency signal of non-infected vital erythrocytes for 48 hours
- Figure 4-13. Experiment design for the observation on infected and non-infected erythrocytes
- Figure 4-14. Comparison of signal changes with different concentrations of parasitemia
- Figure 4-15. Comparison of the measurement of the frequency differential for infected and non-infected erythrocytes
- Figure 4-16. Photographs of the quartz at the end of the experiment
- Figure 4-17. Comparison of signal changes with >80% schizonts haemolysed and non infected erythrocytes
- Figure 4-18. Controls performed to establish the validation between the frequency measurement for the infected erythrocytes and the non-infected erythrocytes
- Figure 4-19. Validation of the frequency measurement for the infected erythrocytes and the non-infected erythrocytes inside the biosensor system
- Figure 4-20. Photographs of Quartz with infected and non-infected erythrocytes
- Figure 4-21. Illustration of reinfection on the quartz in the biosensor system and their external control
-

-
- Figure 4-22. Reinfection of healthy erythrocytes on a second quartz within one QCM system
- Figure 4-23. Apical interaction and tight junction of merozoite on erythrocyte
- Figure 4-24. Photographs of Quartz re-infected
- Figure 4-25. Validation by flow cytometry
- Figure 4-26. Re-infection control
- Figure 4-27. Merozoite viability
- Figure 4-28. Chemical structure of Artesunate
- Figure 4-29. Inhibition with Artesunate of the merozoite release
- Figure 4-30. Validation of Artesunate test
- Figure 4-31. Chemical structure of Chlorotoniil
- Figure 4-32. Inhibition test of Chlorotoniil
- Figure 4-33. Validation of Chlorotoniil test
- Figure 4-34. Chemical structure of the protease inhibitors E64
- Figure 4-35. Inhibition of the merozoite release using protease inhibitors E64
- Figure 4-36. Validation of the test with protease E64
- Figure 4-37. Chemical structure of the protease inhibitors Leupeptin
- Figure 4-38. Inhibition merozoite release using protease inhibitors Leupeptin
- Figure 4-39. Validation test with Leupeptin
- Figure 4-40. Photographs of quartz with exposure of protease inhibitors
-

TABLE OF CONTENTS

	Pages
1. <u>INTRODUCTION AND OBJECTIV OF THE RESEARCH</u>	<u>1</u>
2. <u>THEORY</u>	<u>3</u>
2.1. BIOSENSOR	3
2.1.1. Mechanism of Biosensor	3
2.1.2. Analytical of piezoelectric crystal microbalance	5
2.1.2.1. Piezoelectric effect	5
2.1.2.2. The piezoelectric Crystal	8
2.1.2.3. Relationship between added mass and frequency shift	9
2.1.3. Biosensor in cell biology and pharmacy	10
2.2. MALARIA	12
2.2.1. Epidemiology	12
2.2.2. Biology of parasite	13
2.2.2.1. Invasion of merozoite	16
2.2.2.1.1. Structure of merozoite	16
2.2.2.1.2. Process of invasion	18
2.2.2.2. Release of merozoite	20
2.2.3. Chemotherapy and drug resistance	24
2.2.4. IMMOBILISATION OF ERYTHROCYTES FOR THE ANALYSIS WITH QCM	27
3. <u>METHODS</u>	<u>31</u>
3.1. PREPARATION GOLD SURFACE	31
3.1.1. Cleaning and adhesion of Quartz	31
3.1.2. Coating with Poly-L-Lysine	31
3.2. PREPARATION OF INFECTED ERYTHROCYTES	31
3.2.1. Malaria culture	31
3.2.2. Giemsa staining and parasitemia estimation	31
3.2.3. Dilution of malaria parasites	32
3.2.4. Synchronisation of parasites	32
3.2.5. Separation of schizont stages	33
3.2.6. Sample preparation for use with the biosensor	34
3.3. EXPERIMENT WITH QCM	35
3.3.1. Disinfection of the biosensor system	37
3.3.2. Sterility test for materials and reagents	38
3.3.3. Microscopic control of the quartz	39
3.4. CYTOMETRIC ANALYSE	39

TABLE OF CONTENTS (continued)

	Pages
4. <u>RESULTS AND DISCUSSIONS</u>	40
4.1. CHARACTERISATION AND OPTIMISATION OF THE POLY-L-LYSINE LAYER	40
4.2. OBSERVATION OF THE BIOSENSOR SIGNAL FOR 48 HOURS	51
4.2.1. Sterility	51
4.2.2. Stability of the PLL-erythrocyte layer over 48 hours	52
4.3. OBSERVATION OF INFECTED AND NON-INFECTED ERYTHROCYTES INSIDE THE BIOSENSOR SYSTEM	56
4.3.1. Concentration of the infected cells	56
4.3.2. Handling the erythrocytes infected with <i>P. falciparum</i>	59
4.3.3. Experiment controls	60
4.4. VALIDATION OF THE FREQUENCY MEASUREMENT FOR THE INFECTED ERYTHROCYTES AND THE NON-INFECTED ERYTHROCYTES INSIDE THE BIOSENSOR SYSTEM.	68
4.5. REINFECTION OF THE HEALTHY ERYTHROCYTES INSIDE THE EQUIPMENT	74
4.5.1. Re-infection control and optimisation	82
4.5.1.1. Efficiency of the re-infection method	82
4.5.1.2. Viability of merozoites	85
4.6. STUDY OF THE MEROZOITE INHIBITION	89
4.6.1. Test with Artesunat	89
4.6.2. Test with Chlorotoniil	95
4.6.3. Test with proteases inhibitors	101
5. <u>SUMMARY</u>	115
6. <u>BIBLIOGRAPHY</u>	120
Appendix 1. Relationship Between Added Mass and Frequency Shift	135
Appendix 2. Properties of Some Important Merozoite Proteins in <i>P. falciparum</i>	136
Appendix 3. Clinical correlation of parasitemia	138
Appendix 3. Reagents and Equipments	139
